

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4431

Name: RITTER 1  
Sterilizer ID: V2178198  
04/05/2024 12:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.3 F 0.0  
2:00 162.7 F 2.9  
4:00 179.8 F 6.5  
6:00 202.2 F 11.3  
8:00 217.3 F 4.9  
10:00 228.2 F 9.6  
12:00 234.4 F 10.0  
14:00 240.5 F 11.9  
16:00 239.2 F 10.4  
18:00 251.9 F 16.6  
20:00 262.9 F 23.0  
22:00 272.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.3 F 29.2  
2:00 271.3 F 28.7  
3:00 272.3 F 29.3  
4:00 271.7 F 28.7

Min 271.2 F 28.5  
Max 272.4 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 222.3 F 0.0  
10:00 226.6 F 0.0  
15:00 226.3 F 0.0  
20:00 225.6 F 0.0  
25:00 223.8 F 0.0  
30:00 224.8 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 22:17  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:01:29  
CYCLE COMPLETE

04/05/2024 01:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_