

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000668  
9:58 14/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:03  
STERILIZING 4:14  
136.3 C 230kPa 4:14  
136.3 C 232kPa 4:44  
136.4 C 232kPa 5:14  
136.2 C 231kPa 5:44  
136.4 C 232kPa 6:14  
136.3 C 232kPa 6:44  
136.4 C 233kPa 7:14  
136.4 C 233kPa 7:44  
136.4 C 233kPa 7:44  
VENTING 7:44  
AIR DRYING 8:39  
CYCLE COMPLETE 8:40

-----  
Digital Signature #  
29EB0ED0D2C3CAC6  
-----