

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3835

Name: RITTER 1  
Sterilizer ID: V2178198  
09/28/2023 12:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
2:00 190.0 F 3.5  
4:00 205.6 F 8.8  
6:00 224.3 F 7.8  
8:00 233.6 F 10.1  
10:00 235.5 F 7.7  
12:00 244.8 F 12.5  
14:00 245.3 F 12.7  
16:00 258.2 F 19.5  
18:00 269.5 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.2 F 28.2  
2:00 272.5 F 28.7  
3:00 271.6 F 28.3  
4:00 272.0 F 28.3  
  
Min 271.4 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 230.1 F 0.0  
10:00 206.9 F 0.0  
15:00 207.1 F 0.0  
20:00 224.4 F 0.0  
25:00 237.1 F 0.0  
30:00 231.7 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:07  
CYCLE COMPLETE

09/28/2023 01:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_