

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2092

Name: RITTER2  
Sterilizer ID: V2313228  
02/27/2023 10:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.2 F 0.0  
2:00 184.3 F 3.1  
4:00 191.9 F 7.4  
6:00 207.7 F 12.8  
8:00 226.8 F 8.1  
10:00 236.3 F 11.9  
12:00 239.5 F 11.7  
14:00 239.1 F 10.5  
16:00 252.8 F 17.4  
18:00 264.6 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.4 F 30.0  
2:00 272.2 F 29.4  
3:00 271.3 F 29.0  
4:00 271.8 F 29.8  
  
Min 271.2 F 28.9  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
5:00 221.3 F 0.0  
10:00 225.0 F 0.0  
15:00 229.0 F 0.0  
20:00 220.9 F 0.0  
25:00 219.6 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 19:54  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

02/27/2023 11:56 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_