
Midmark M11 - v1.0.0
Total Cycles: 1362

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2021 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.0 F 0.0
2:00 153.2 F 1.9
4:00 171.1 F 5.4
6:00 196.9 F 10.5
8:00 213.6 F 4.6
10:00 228.6 F 10.0
12:00 232.9 F 8.1
14:00 242.5 F 12.4
16:00 241.1 F 10.8
18:00 254.4 F 17.6
20:00 266.0 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.8 F 29.0
2:00 271.6 F 28.3
3:00 272.8 F 29.2
4:00 272.3 F 28.7

Min 271.6 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 223.4 F 0.0
10:00 227.0 F 0.0
15:00 227.5 F 0.0
20:00 225.8 F 0.0
25:00 227.4 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

11/11/2021 01:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____