

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1521

Name: RITTER2  
Sterilizer ID: V2313228  
09/22/2022 08:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.6 F 0.0  
2:00 157.3 F 1.6  
4:00 160.1 F 5.1  
6:00 189.0 F 10.2  
8:00 220.1 F 5.1  
10:00 231.9 F 9.8  
12:00 234.2 F 8.7  
14:00 242.5 F 12.6  
16:00 240.8 F 11.3  
18:00 253.6 F 17.8  
20:00 265.2 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.9 F 30.0  
2:00 272.2 F 29.5  
3:00 271.1 F 29.0  
4:00 272.1 F 29.9  
  
Min 271.0 F 29.0  
Max 273.1 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 233.6 F 0.0  
10:00 220.7 F 0.0  
15:00 231.4 F 0.0  
20:00 228.1 F 0.0  
25:00 231.8 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 21:48  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:03  
CYCLE COMPLETE

09/22/2022 09:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_