
Midmark M11 - v1.0.5
Total Cycles: 2967

Name: RITTER2
Sterilizer ID: V2313228
10/04/2023 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.9 F 0.0
2:00 155.5 F 2.0
4:00 173.6 F 5.3
6:00 193.1 F 9.7
8:00 216.5 F 7.4
10:00 223.1 F 8.6
12:00 239.1 F 12.6
14:00 240.4 F 11.2
16:00 238.1 F 9.8
18:00 250.9 F 16.2
20:00 262.6 F 22.8
22:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 271.6 F 29.0
2:00 271.8 F 29.3
3:00 271.8 F 29.7
4:00 272.0 F 29.6

Min 270.9 F 28.9
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 226.9 F 0.0
10:00 215.6 F 0.0
15:00 182.7 F 0.0
20:00 177.7 F 0.0
25:00 174.8 F 0.0
30:00 175.1 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:20
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

10/04/2023 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____