
Midmark M11 - v1.0.0
Total Cycles: 2647

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2022 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 222.5 F 0.0
2:00 215.8 F 3.4
4:00 231.3 F 9.3
6:00 234.8 F 7.3
8:00 240.4 F 12.2
10:00 239.4 F 10.2
12:00 237.8 F 8.9
14:00 252.8 F 16.5
16:00 265.4 F 24.0
18:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.6 F 29.2
2:00 272.0 F 29.1
3:00 272.6 F 29.1
4:00 271.9 F 29.1

Min 271.6 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.0 F 0.0
5:00 230.2 F 0.0
10:00 224.4 F 0.0
15:00 227.5 F 0.0
20:00 229.8 F 0.0
25:00 228.0 F 0.0
30:00 229.2 F 0.0

Drying Complete

FILLING: 2:42
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

12/06/2022 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____