
Midmark M11 - v1.0.5
Total Cycles: 3750

Name: RITTER2
Sterilizer ID: V2313228
07/03/2024 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.3 F 0.0
2:00 69.5 F 0.0
4:00 74.6 F 0.1
6:00 104.2 F 0.3
8:00 143.2 F 3.0
10:00 169.2 F 5.8
12:00 194.0 F 9.1
14:00 205.2 F 13.4
16:00 215.9 F 6.2
18:00 227.1 F 10.3
20:00 231.7 F 8.8
22:00 239.4 F 11.6
24:00 237.5 F 9.8
26:00 250.6 F 16.0
28:00 261.6 F 22.4
30:00 271.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.0
1:00 272.7 F 29.8
2:00 271.7 F 28.9
3:00 272.8 F 29.8
4:00 271.7 F 28.9

Min 271.3 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 215.0 F 0.0
10:00 189.1 F 0.0
15:00 186.2 F 0.0
20:00 184.7 F 0.0
25:00 200.2 F 0.0
30:00 201.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:08:37
CYCLE COMPLETE

07/03/2024 09:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____