
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007147
7:18 05/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:41
STERILIZING 3:49
136.4 C 220kPa 3:49
136.3 C 223kPa 4:19
136.4 C 223kPa 4:49
136.2 C 221kPa 5:19
136.4 C 218kPa 5:49
136.3 C 220kPa 6:19
136.4 C 222kPa 6:49
136.4 C 221kPa 7:19
136.4 C 221kPa 7:19
VENTING 7:19
AIR DRYING 8:11
CYCLE COMPLETE 38:12

Digital Signature #
B915A34FBA766DB1
