

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1637

Name: RITTER2  
Sterilizer ID: V2313228  
10/21/2022 11:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.2 F 0.0  
2:00 133.3 F 2.0  
4:00 153.6 F 5.2  
6:00 180.9 F 10.0  
8:00 206.8 F 4.5  
10:00 222.5 F 8.6  
12:00 233.8 F 11.6  
14:00 239.0 F 11.4  
16:00 238.5 F 10.2  
18:00 252.4 F 17.0  
20:00 264.0 F 23.9  
22:00 271.6 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.7 F 29.6  
2:00 271.7 F 29.0  
3:00 273.2 F 29.7  
4:00 272.0 F 29.0

Min 271.6 F 28.9  
Max 273.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 203.8 F 0.0  
10:00 217.0 F 0.0  
15:00 220.9 F 0.0  
20:00 223.7 F 0.0  
25:00 225.7 F 0.0  
30:00 224.6 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 22:01  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:14  
CYCLE COMPLETE

10/21/2022 12:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_