
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002416
9:34 13/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:01
STERILIZING 4:03
136.5 C 225kPa 4:03
136.3 C 224kPa 4:33
136.1 C 222kPa 5:03
136.4 C 225kPa 5:33
136.4 C 225kPa 6:03
136.2 C 223kPa 6:33
136.3 C 224kPa 7:03
136.3 C 223kPa 7:33
136.3 C 223kPa 7:33
VENTING 7:33
AIR DRYING 8:27
CYCLE COMPLETE 21:52

Digital Signature #
A8F6D3FF254C6607
