
Midmark M11 - v1.0.0
Total Cycles: 2918

Name: RITTER 1
Sterilizer ID: V2178198
02/17/2023 11:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 96.2 F 0.0
4:00 88.8 F 0.0
6:00 116.9 F 0.1
8:00 150.8 F 0.4
10:00 177.4 F 2.0
12:00 192.6 F 4.9
14:00 217.5 F 7.8
16:00 224.1 F 11.3
18:00 224.8 F 5.6
20:00 235.1 F 10.0
22:00 235.3 F 9.0
24:00 244.0 F 12.4
26:00 240.8 F 10.8
28:00 253.5 F 17.0
30:00 264.6 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.7
2:00 271.6 F 28.6
3:00 272.1 F 28.4
4:00 272.0 F 28.9

Min 271.0 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 212.5 F 0.0
10:00 220.2 F 0.0
15:00 235.7 F 0.0
20:00 234.4 F 0.0
25:00 233.1 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:58
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:10:17
CYCLE COMPLETE

02/17/2023 12:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____