
Midmark M11 - v1.0.5
Total Cycles: 1118

Name: RITTER2
Sterilizer ID: V2313228
05/26/2022 06:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 77.1 F 0.0
4:00 78.9 F 0.0
6:00 107.1 F 0.1
8:00 160.7 F 1.5
10:00 194.6 F 5.4
12:00 197.9 F 9.1
14:00 219.2 F 12.9
16:00 225.4 F 6.3
18:00 234.4 F 10.4
20:00 234.6 F 8.8
22:00 242.6 F 12.3
24:00 240.3 F 10.8
26:00 253.0 F 17.1
28:00 263.8 F 23.5
30:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.0 F 29.1
2:00 272.7 F 29.5
3:00 272.0 F 29.6
4:00 271.8 F 29.0

Min 271.5 F 28.9
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 211.3 F 0.0
10:00 211.8 F 0.0
15:00 206.9 F 0.0
20:00 196.3 F 0.0
Dry Cycle Aborted

FILLING: 0:57
HEATING: 30:47
STERILIZING: 4:00
VENTING: 3:01
DRYING: 20:06
TOTAL CYCLE: 00:58:51

05/26/2022 07:48 AM
ACCEPTED: X
REJECTED:
Operator: U02