
Midmark M11 - v1.0.0
Total Cycles: 627

Name: MIDMARK
Sterilizer ID: V2178198
01/26/2021 07:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 77.6 F 0.0
4:00 124.9 F 0.0
6:00 160.4 F 0.1
8:00 165.8 F 0.3
10:00 194.6 F 1.6
12:00 210.5 F 5.1
14:00 222.1 F 8.1
16:00 232.3 F 11.8
18:00 231.0 F 7.3
20:00 242.6 F 12.7
22:00 240.2 F 10.4
24:00 236.1 F 8.4
26:00 251.8 F 15.8
28:00 264.0 F 22.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.0 F 28.1
2:00 272.6 F 28.6
3:00 271.5 F 28.0
4:00 272.0 F 28.2

Min 271.3 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 211.1 F 0.0
10:00 240.2 F 0.0
Dry Cycle Aborted

FILLING: 0:50
HEATING: 29:55
STERILIZING: 4:00
VENTING: 2:51
DRYING: 14:10
TOTAL CYCLE: 00:51:46

01/26/2021 08:16 AM
ACCEPTED: X
REJECTED:
Operator: U01