

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3274

Name: RITTER2  
Sterilizer ID: V2313228  
01/04/2024 09:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 70.6 F 0.0  
4:00 73.6 F 0.0  
6:00 91.9 F 0.6  
8:00 128.7 F 2.7  
10:00 154.6 F 5.7  
12:00 175.8 F 9.2  
14:00 193.9 F 12.8  
16:00 214.5 F 5.2  
18:00 227.9 F 8.8  
20:00 237.2 F 13.0  
22:00 235.7 F 9.4  
24:00 239.9 F 10.4  
26:00 244.9 F 13.0  
28:00 256.4 F 18.9  
30:00 266.6 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.8  
1:00 272.1 F 29.3  
2:00 271.7 F 29.2  
3:00 272.7 F 29.5  
4:00 272.6 F 29.9

Min 271.0 F 28.6  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.1 F 0.0  
5:00 237.6 F 0.0  
10:00 224.6 F 0.0  
15:00 212.9 F 0.0  
20:00 188.5 F 0.0  
25:00 186.3 F 0.0  
30:00 183.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:40  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:09:47  
CYCLE COMPLETE

01/04/2024 10:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_