

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008348  
11:09 25/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:12  
STERILIZING 4:18  
136.0 C 220kPa 4:18  
136.1 C 223kPa 4:48  
136.3 C 223kPa 5:18  
135.9 C 222kPa 5:48  
136.4 C 224kPa 6:18  
136.4 C 224kPa 6:48  
136.3 C 221kPa 7:18  
136.3 C 221kPa 7:48  
136.3 C 221kPa 7:48  
VENTING 7:48  
AIR DRYING 8:44  
CYCLE COMPLETE 17:45

-----

Digital Signature #  
D53505E2DDB2E96C  
-----