

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3598

Name: RITTER2  
Sterilizer ID: V2313228  
05/08/2024 11:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.5 F 0.0  
2:00 114.6 F 0.8  
4:00 128.9 F 3.0  
6:00 149.9 F 6.0  
8:00 175.7 F 9.8  
10:00 202.3 F 13.0  
12:00 213.8 F 6.6  
14:00 224.3 F 10.8  
16:00 228.0 F 7.4  
18:00 240.1 F 12.4  
20:00 239.3 F 10.7  
22:00 251.2 F 16.7  
24:00 261.8 F 23.0  
26:00 272.1 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.8  
1:00 273.3 F 30.2  
2:00 271.9 F 30.2  
3:00 272.7 F 30.1  
4:00 272.0 F 30.2

Min 271.5 F 29.7  
Max 273.5 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 191.2 F 0.0  
10:00 194.1 F 0.0  
15:00 192.3 F 0.0  
20:00 189.3 F 0.0  
25:00 189.1 F 0.0  
30:00 188.8 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 27:20  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:05:49  
CYCLE COMPLETE

05/08/2024 12:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_