
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004048
8:15 19/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:07
STERILIZING 4:15
136.0 C 220kPa 4:15
136.2 C 220kPa 4:45
135.9 C 220kPa 5:15
136.4 C 224kPa 5:45
136.3 C 222kPa 6:15
136.4 C 221kPa 6:45
136.2 C 221kPa 7:15
136.2 C 220kPa 7:45
136.2 C 220kPa 7:45
VENTING 7:45
AIR DRYING 8:34
CYCLE COMPLETE 9:11

Digital Signature #
3EB1648FC5B81518
