

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1358

Name: RITTER2  
Sterilizer ID: V2313228  
07/27/2022 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 181.9 F 0.2  
4:00 205.5 F 3.8  
6:00 218.7 F 8.8  
8:00 226.1 F 10.9  
10:00 232.0 F 10.5  
12:00 232.3 F 8.1  
14:00 239.7 F 10.8  
16:00 246.9 F 14.3  
18:00 259.6 F 21.2  
20:00 270.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.7 F 29.8  
2:00 272.0 F 29.1  
3:00 271.9 F 29.4  
4:00 272.6 F 29.8

Min 271.5 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.8 F 0.0  
5:00 236.0 F 0.0  
10:00 233.1 F 0.0  
15:00 230.9 F 0.0  
20:00 228.5 F 0.0  
25:00 230.7 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 20:48  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:07  
CYCLE COMPLETE

07/27/2022 03:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_