
Midmark M11 - v1.0.5
Total Cycles: 3100

Name: RITTER2
Sterilizer ID: V2313228
11/08/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.4 F 0.0
2:00 125.7 F 0.8
4:00 144.2 F 3.5
6:00 168.9 F 6.9
8:00 190.1 F 10.9
10:00 220.9 F 6.0
12:00 218.4 F 9.9
14:00 231.5 F 9.3
16:00 236.9 F 11.0
18:00 234.6 F 8.8
20:00 248.3 F 15.0
22:00 259.4 F 21.1
24:00 269.5 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 271.5 F 29.6
2:00 271.5 F 29.5
3:00 271.8 F 29.6
4:00 272.4 F 29.8

Min 271.3 F 29.5
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 201.7 F 0.0
10:00 191.6 F 0.0
15:00 192.5 F 0.0
20:00 192.1 F 0.0
25:00 191.5 F 0.0
30:00 191.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:44
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:20
CYCLE COMPLETE

11/08/2023 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____