
Midmark M11 - v1.0.5
Total Cycles: 3833

Name: RITTER2
Sterilizer ID: V2313228
08/07/2024 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.9 F 0.0
2:00 118.1 F 0.6
4:00 124.9 F 2.7
6:00 160.3 F 6.0
8:00 190.2 F 9.1
10:00 199.6 F 12.9
12:00 216.2 F 6.1
14:00 225.3 F 9.9
16:00 234.0 F 10.4
18:00 236.8 F 10.8
20:00 237.1 F 8.4
22:00 247.2 F 14.4
24:00 257.5 F 20.1
26:00 267.5 F 26.3
28:00 272.9 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.8
1:00 272.5 F 30.1
2:00 272.5 F 29.8
3:00 272.0 F 29.8
4:00 272.1 F 30.2

Min 271.9 F 29.7
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 217.4 F 0.0
10:00 219.4 F 0.0
15:00 224.5 F 0.0
20:00 228.0 F 0.0
25:00 231.9 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:10
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:52
CYCLE COMPLETE

08/07/2024 12:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____