
Midmark M11 - v1.0.5
Total Cycles: 1473

Name: RITTER2
Sterilizer ID: V2313228
09/12/2022 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 160.1 F 1.7
4:00 161.4 F 4.6
6:00 199.9 F 9.2
8:00 215.3 F 8.2
10:00 226.9 F 10.9
12:00 230.7 F 7.5
14:00 243.6 F 13.1
16:00 242.4 F 12.1
18:00 254.5 F 18.4
20:00 265.7 F 25.1
22:00 271.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.7
1:00 271.2 F 29.9
2:00 271.4 F 30.1
3:00 271.5 F 30.1
4:00 271.3 F 30.0

Min 271.1 F 29.6
Max 272.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 229.8 F 0.0
10:00 237.6 F 0.0
15:00 228.9 F 0.0
20:00 226.2 F 0.0
25:00 230.0 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:26
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

09/12/2022 09:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____