
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001089
12:59 12/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:19
STERILIZING 4:21
136.4 C 224kPa 4:21
136.3 C 223kPa 4:51
136.4 C 224kPa 5:21
136.4 C 224kPa 5:51
136.4 C 224kPa 6:21
136.4 C 225kPa 6:51
136.4 C 224kPa 7:21
136.3 C 220kPa 7:51
136.3 C 220kPa 7:51
VENTING 7:51
AIR DRYING 8:46
CYCLE COMPLETE 8:49

Digital Signature #
171269A2C006E7EF
