

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1696

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2022 09:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.4 F 0.0  
2:00 151.1 F 0.0  
4:00 161.4 F 0.3  
6:00 205.7 F 1.2  
8:00 228.4 F 5.6  
10:00 237.7 F 10.5  
12:00 233.0 F 7.6  
14:00 245.1 F 12.9  
16:00 241.9 F 11.0  
18:00 239.0 F 9.7  
20:00 252.6 F 16.4  
22:00 263.7 F 23.2  
24:00 270.5 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.8 F 29.0  
1:00 271.5 F 29.1  
2:00 272.1 F 29.3  
3:00 272.4 F 29.1  
4:00 272.2 F 29.0

Min 271.3 F 28.9  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.6 F 0.0  
5:00 233.8 F 0.0  
10:00 229.9 F 0.0  
15:00 230.7 F 0.0  
20:00 226.3 F 0.0  
25:00 228.3 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:37  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:41  
CYCLE COMPLETE

11/08/2022 10:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_