

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1134

Name: RITTER2  
Sterilizer ID: V2313228  
05/30/2022 07:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.6 F 0.0  
2:00 81.4 F 0.0  
4:00 85.8 F 0.0  
6:00 126.6 F 0.5  
8:00 165.1 F 3.5  
10:00 188.1 F 6.9  
12:00 213.2 F 10.5  
14:00 221.2 F 6.4  
16:00 225.9 F 7.5  
18:00 236.3 F 11.5  
20:00 231.3 F 7.4  
22:00 243.4 F 12.9  
24:00 240.5 F 11.0  
26:00 252.6 F 17.0  
28:00 263.1 F 23.2  
30:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 271.5 F 29.2  
2:00 271.7 F 29.2  
3:00 272.4 F 29.3  
4:00 272.8 F 29.6

Min 271.4 F 29.2  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.7 F 0.0  
5:00 225.3 F 0.0  
10:00 221.5 F 0.0  
15:00 231.4 F 0.0  
20:00 212.7 F 0.0  
25:00 222.7 F 0.0  
30:00 190.7 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:05  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:08:57  
CYCLE COMPLETE

05/30/2022 08:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_