

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3008

Name: RITTER 1  
Sterilizer ID: V2178198  
03/15/2023 10:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
2:00 202.0 F 4.3  
4:00 216.1 F 9.3  
6:00 225.1 F 7.2  
8:00 236.2 F 10.1  
10:00 233.4 F 7.9  
12:00 244.6 F 12.4  
14:00 245.4 F 12.8  
16:00 258.7 F 19.9  
18:00 270.4 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 29.1  
2:00 272.1 F 28.7  
3:00 271.5 F 28.2  
4:00 271.3 F 28.4  
  
Min 271.2 F 28.1  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.7 F 0.0  
5:00 225.6 F 0.0  
10:00 236.5 F 0.0  
15:00 239.3 F 0.0  
20:00 230.6 F 0.0  
25:00 223.0 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 18:54  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:47  
CYCLE COMPLETE

03/15/2023 11:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_