

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001490  
10:31 08/04/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:42  
PRESSURIZING 2:47  
STERILIZING 3:54  
136.5 C 220kPa 3:54  
136.3 C 224kPa 4:24  
136.4 C 221kPa 4:54  
136.4 C 221kPa 5:24  
136.4 C 224kPa 5:54  
136.2 C 223kPa 6:24  
136.2 C 221kPa 6:54  
136.4 C 224kPa 7:24  
136.4 C 224kPa 7:24  
VENTING 7:24  
AIR DRYING 8:19  
CYCLE COMPLETE 8:26

-----  
Digital Signature #  
EE6F12932A204BF7  
-----