

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1171

Name: RITTER 1  
Sterilizer ID: V2178198  
09/14/2021 02:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
2:00 180.5 F 2.4  
4:00 181.3 F 5.8  
6:00 218.3 F 9.8  
8:00 225.5 F 6.8  
10:00 234.6 F 9.7  
12:00 237.8 F 9.5  
14:00 243.0 F 12.0  
16:00 240.0 F 10.3  
18:00 252.8 F 16.6  
20:00 263.8 F 23.1  
22:00 271.5 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.2 F 28.6  
2:00 272.7 F 28.9  
3:00 272.4 F 29.1  
4:00 271.3 F 28.6

Min 271.3 F 28.5  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 238.4 F 0.0  
10:00 226.6 F 0.0  
15:00 222.8 F 0.0  
20:00 224.0 F 0.0  
25:00 227.8 F 0.0  
30:00 230.0 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 22:47  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:01:53  
CYCLE COMPLETE

09/14/2021 03:07 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01