

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3629

Name: RITTER2  
Sterilizer ID: V2313228  
05/21/2024 08:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.1 F 0.0  
2:00 71.4 F 0.0  
4:00 106.2 F 0.0  
6:00 96.8 F 0.4  
8:00 131.6 F 2.5  
10:00 156.2 F 5.6  
12:00 174.3 F 9.1  
14:00 191.6 F 12.6  
16:00 211.3 F 4.6  
18:00 221.0 F 8.0  
20:00 231.1 F 11.7  
22:00 230.9 F 7.9  
24:00 241.7 F 12.9  
26:00 240.4 F 11.0  
28:00 251.8 F 16.7  
30:00 261.9 F 22.6  
32:00 271.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.2 F 29.3  
2:00 272.2 F 29.7  
3:00 271.7 F 28.9  
4:00 272.4 F 29.6  
  
Min 271.5 F 28.9  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
5:00 220.4 F 0.0  
10:00 195.9 F 0.0  
15:00 191.1 F 0.0  
20:00 187.4 F 0.0  
25:00 187.8 F 0.0  
30:00 195.1 F 0.0  
  
Drying Complete

FILLING: 0:58  
HEATING: 32:36  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:10:55  
CYCLE COMPLETE

05/21/2024 10:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_