
Midmark M11 - v1.0.0
Total Cycles: 5476

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2025 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 165.6 F 3.4
4:00 181.4 F 7.5
6:00 203.3 F 12.4
8:00 220.8 F 5.7
10:00 233.9 F 10.6
12:00 232.8 F 7.5
14:00 244.0 F 13.0
16:00 242.0 F 11.5
18:00 254.5 F 17.9
20:00 265.3 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.3
2:00 271.5 F 28.5
3:00 271.9 F 29.2
4:00 273.0 F 29.7

Min 271.0 F 28.4
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 221.0 F 0.0
10:00 194.9 F 0.0
15:00 182.1 F 0.0
20:00 178.1 F 0.0
25:00 176.3 F 0.0
30:00 175.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

04/03/2025 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____