
Midmark M11 - v1.0.5
Total Cycles: 3603

Name: RITTER2
Sterilizer ID: V2313228
05/09/2024 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.8 F 0.0
2:00 121.4 F 1.3
4:00 146.3 F 4.0
6:00 171.4 F 7.7
8:00 210.2 F 11.9
10:00 214.9 F 5.6
12:00 224.3 F 10.3
14:00 230.2 F 7.9
16:00 239.1 F 11.3
18:00 235.8 F 9.0
20:00 249.3 F 15.2
22:00 260.4 F 21.5
24:00 270.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.0 F 29.8
2:00 272.4 F 29.4
3:00 272.6 F 30.1
4:00 272.4 F 29.4

Min 270.9 F 28.8
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 228.1 F 0.0
10:00 188.9 F 0.0
15:00 186.3 F 0.0
20:00 183.2 F 0.0
25:00 182.1 F 0.0
30:00 181.5 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:40
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:60
CYCLE COMPLETE

05/09/2024 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____