
Midmark M11 - v1.0.0
Total Cycles: 2585

Name: RITTER 1
Sterilizer ID: V2178198
11/21/2022 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.3 F 0.0
2:00 198.9 F 4.2
4:00 200.1 F 9.3
6:00 220.1 F 6.8
8:00 231.3 F 10.5
10:00 231.9 F 7.6
12:00 241.3 F 11.0
14:00 246.5 F 13.5
16:00 259.3 F 20.5
18:00 270.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.4
3:00 272.3 F 29.2
4:00 272.5 F 28.9

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 230.6 F 0.0
10:00 247.4 F 0.0
15:00 240.1 F 0.0
20:00 228.6 F 0.0
25:00 223.8 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

11/21/2022 01:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____