
Midmark M11 - v1.0.5
Total Cycles: 2655

Name: RITTER2
Sterilizer ID: V2313228
07/17/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 181.6 F 3.0
4:00 197.4 F 7.5
6:00 215.0 F 12.3
8:00 224.5 F 7.2
10:00 237.7 F 12.2
12:00 237.7 F 10.3
14:00 237.1 F 8.9
16:00 250.6 F 16.0
18:00 262.1 F 22.7
20:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 28.8
3:00 272.4 F 29.7
4:00 272.0 F 29.1

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 232.6 F 0.0
10:00 222.6 F 0.0
15:00 212.2 F 0.0
20:00 205.7 F 0.0
25:00 199.4 F 0.0
30:00 198.3 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

07/17/2023 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____