
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002763
15:56 30/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:29
STERILIZING 4:25
136.0 C 220kPa 4:25
136.1 C 220kPa 4:55
136.4 C 222kPa 5:25
136.4 C 225kPa 5:55
136.1 C 222kPa 6:25
136.4 C 222kPa 6:55
136.4 C 222kPa 7:25
136.4 C 225kPa 7:55
136.4 C 225kPa 7:55
VENTING 7:55
AIR DRYING 8:51
CYCLE COMPLETE 9:13

Digital Signature #
587BAF7853E888D0
