
Midmark M11 - v1.0.5
Total Cycles: 4277

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C232: FILL MODE
WATER LOW
CHAMBER TEMP 69F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

01/22/2025 09:18 AM