

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1158

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2022 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.6 F 0.0  
2:00 128.5 F 0.5  
4:00 141.2 F 3.2  
6:00 168.6 F 7.0  
8:00 196.2 F 11.6  
10:00 216.3 F 5.5  
12:00 227.3 F 10.3  
14:00 231.6 F 8.1  
16:00 240.0 F 12.3  
18:00 239.4 F 10.6  
20:00 251.8 F 16.9  
22:00 263.0 F 23.4  
24:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.5 F 29.9  
2:00 271.6 F 29.1  
3:00 272.6 F 30.0  
4:00 272.1 F 29.4

Min 271.3 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 231.2 F 0.0  
10:00 222.0 F 0.0  
15:00 224.8 F 0.0  
20:00 224.7 F 0.0  
25:00 218.0 F 0.0  
30:00 221.0 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:21  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:02:22  
CYCLE COMPLETE

06/06/2022 11:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_