

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4738

Name: RITTER 1  
Sterilizer ID: V2178198  
07/16/2024 05:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.4 F 0.0  
2:00 182.7 F 4.2  
4:00 207.5 F 9.4  
6:00 221.5 F 7.5  
8:00 228.0 F 10.4  
10:00 231.5 F 7.4  
12:00 242.4 F 11.8  
14:00 244.8 F 13.0  
16:00 257.7 F 19.8  
18:00 269.1 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.2 F 28.9  
2:00 272.5 F 29.3  
3:00 271.4 F 28.8  
4:00 272.3 F 29.0  
  
Min 271.3 F 28.6  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 232.5 F 0.0  
10:00 226.7 F 0.0  
15:00 226.6 F 0.0  
20:00 226.4 F 0.0  
25:00 228.6 F 0.0  
30:00 225.6 F 0.0  
  
Drying Complete

FILLING: 1:57  
HEATING: 19:01  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 00:58:30  
CYCLE COMPLETE

07/16/2024 06:37 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_