

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5071

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2024 06:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
2:00 194.8 F 2.8  
4:00 210.6 F 8.8  
6:00 223.2 F 8.0  
8:00 232.8 F 9.9  
10:00 234.4 F 8.0  
12:00 245.0 F 12.5  
14:00 245.4 F 12.8  
16:00 258.8 F 19.9  
18:00 270.5 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.1 F 28.3  
2:00 272.7 F 28.8  
3:00 271.7 F 28.3  
4:00 272.1 F 28.4  
  
Min 271.5 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 223.5 F 0.0  
10:00 229.6 F 0.0  
15:00 228.9 F 0.0  
20:00 229.4 F 0.0  
25:00 230.7 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 1:45  
HEATING: 18:43  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

11/07/2024 07:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_