
Midmark M11 - v1.0.0
Total Cycles: 1405

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 152.3 F 1.5
4:00 166.3 F 4.7
6:00 189.7 F 8.8
8:00 212.8 F 9.7
10:00 222.9 F 7.9
12:00 234.3 F 12.7
14:00 235.3 F 9.2
16:00 238.8 F 10.0
18:00 246.0 F 13.2
20:00 258.0 F 19.7
22:00 268.9 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.2
2:00 272.1 F 28.5
3:00 271.4 F 28.5
4:00 272.7 F 29.2

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 205.7 F 0.0
10:00 216.6 F 0.0
15:00 215.6 F 0.0
20:00 218.1 F 0.0
25:00 222.0 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

11/30/2021 10:25 AM
ACCEPTED: X
REJECTED:
Operator: U02