

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2392

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2022 07:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.4 F 0.0  
2:00 82.9 F 0.0  
4:00 96.5 F 0.1  
6:00 131.6 F 1.5  
8:00 153.0 F 3.2  
10:00 173.9 F 5.9  
12:00 193.5 F 9.5  
14:00 210.9 F 13.4  
16:00 221.1 F 5.7  
18:00 230.4 F 9.7  
20:00 236.7 F 10.3  
22:00 239.9 F 11.1  
24:00 236.6 F 8.8  
26:00 250.5 F 15.4  
28:00 261.8 F 21.9  
30:00 272.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.7 F 28.9  
2:00 271.5 F 28.2  
3:00 272.7 F 28.9  
4:00 271.6 F 28.1  
  
Min 271.5 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
5:00 202.6 F 0.0  
10:00 201.7 F 0.0  
15:00 199.7 F 0.0  
20:00 220.9 F 0.0  
25:00 226.5 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 30:28  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:22  
CYCLE COMPLETE

10/04/2022 08:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_