
Midmark M11 - v1.0.0
Total Cycles: 4053

Name: RITTER 1
Sterilizer ID: V2178198
11/23/2023 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 75.6 F 0.0
4:00 107.4 F 0.4
6:00 145.5 F 2.4
8:00 177.2 F 4.9
10:00 196.9 F 7.7
12:00 210.2 F 11.5
14:00 212.7 F 2.7
16:00 221.4 F 5.7
18:00 229.7 F 9.0
20:00 237.7 F 12.8
22:00 233.1 F 8.6
24:00 243.8 F 13.5
26:00 242.2 F 11.5
28:00 253.8 F 17.4
30:00 264.1 F 23.6
32:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.8 F 29.2
2:00 271.8 F 28.4
3:00 272.3 F 29.1
4:00 272.5 F 28.9

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 210.5 F 0.0
10:00 213.2 F 0.0
15:00 217.7 F 0.0
20:00 216.9 F 0.0
25:00 218.7 F 0.0
30:00 220.1 F 0.0

Drying Complete

FILLING: 0:59
HEATING: 32:10
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:10:42
CYCLE COMPLETE

11/23/2023 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____