

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3716

Name: RITTER 1  
Sterilizer ID: V2178198  
09/05/2023 10:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
2:00 209.2 F 1.9  
4:00 205.6 F 6.5  
6:00 221.3 F 12.0  
8:00 228.3 F 7.7  
10:00 240.7 F 13.3  
12:00 241.0 F 11.3  
14:00 239.3 F 10.0  
16:00 253.3 F 16.9  
18:00 264.9 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.4 F 28.6  
2:00 271.8 F 28.6  
3:00 272.2 F 28.5  
4:00 271.7 F 28.5  
  
Min 271.4 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 229.3 F 0.0  
10:00 220.0 F 0.0  
15:00 214.8 F 0.0  
20:00 209.7 F 0.0  
25:00 211.5 F 0.0  
30:00 207.4 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

09/05/2023 11:22 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_