

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4558

Name: RITTER 1  
Sterilizer ID: V2178198  
05/15/2024 01:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.9 F 0.0  
2:00 163.3 F 3.0  
4:00 178.0 F 6.4  
6:00 195.9 F 10.4  
8:00 220.5 F 5.4  
10:00 228.6 F 8.0  
12:00 233.7 F 12.5  
14:00 236.0 F 9.1  
16:00 239.6 F 10.6  
18:00 245.6 F 13.0  
20:00 257.0 F 19.0  
22:00 267.3 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.3  
1:00 272.6 F 28.8  
2:00 271.0 F 28.2  
3:00 272.5 F 29.4  
4:00 272.8 F 29.1

Min 271.0 F 28.2  
Max 273.2 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
5:00 198.7 F 0.0  
10:00 195.1 F 0.0  
15:00 193.0 F 0.0  
20:00 192.9 F 0.0  
25:00 202.2 F 0.0  
30:00 209.8 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 23:31  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:02:39  
CYCLE COMPLETE

05/15/2024 02:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_