
Midmark M11 - v1.0.5
Total Cycles: 3136

Name: RITTER2
Sterilizer ID: V2313228
11/16/2023 04:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.5 F 0.0
2:00 166.5 F 3.0
4:00 186.6 F 7.4
6:00 206.7 F 12.5
8:00 221.7 F 7.3
10:00 231.3 F 12.7
12:00 236.9 F 10.4
14:00 236.1 F 9.2
16:00 251.1 F 16.4
18:00 263.5 F 23.6
20:00 271.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.5
2:00 271.3 F 29.0
3:00 272.5 F 29.6
4:00 271.7 F 29.0

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.0 F 0.0
5:00 226.3 F 0.0
10:00 198.7 F 0.0
15:00 193.3 F 0.0
20:00 190.1 F 0.0
25:00 189.4 F 0.0
30:00 188.0 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

11/16/2023 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____