
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000202
12:11 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:45
STERILIZING 5:22
136.0 C 228kPa 5:22
136.2 C 232kPa 5:52
136.4 C 233kPa 6:22
136.3 C 232kPa 6:52
136.3 C 232kPa 7:22
136.2 C 232kPa 7:52
136.3 C 233kPa 8:22
136.2 C 232kPa 8:52
136.2 C 232kPa 8:52
VENTING 8:52
AIR DRYING 9:54
CYCLE COMPLETE 10:25

Digital Signature #
A5E460BE55A4A193
