

-----  
Midmark M11 - v1.0.0  
Total Cycles: 810

Name: RITTER 1  
Sterilizer ID: V2178198  
04/26/2021 01:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.8 F 0.0  
2:00 179.2 F 3.5  
4:00 205.4 F 9.1  
6:00 218.4 F 7.5  
8:00 232.6 F 10.0  
10:00 233.9 F 8.2  
12:00 244.2 F 12.2  
14:00 246.0 F 13.2  
16:00 259.3 F 20.4  
18:00 271.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.4 F 28.5  
2:00 272.7 F 29.0  
3:00 271.9 F 28.2  
4:00 273.0 F 29.0  
  
Min 271.7 F 28.1  
Max 273.0 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.0 F 0.0  
5:00 217.0 F 0.0  
10:00 246.9 F 0.0  
15:00 237.8 F 0.0  
20:00 220.3 F 0.0  
25:00 210.6 F 0.0  
30:00 203.2 F 0.0  
Drying Complete

FILLING: 2:16  
HEATING: 18:40  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:04  
CYCLE COMPLETE

04/26/2021 02:57 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01