
Midmark M11 - v1.0.5
Total Cycles: 1426

Name: RITTER2
Sterilizer ID: V2313228
08/29/2022 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.1 F 0.0
2:00 136.5 F 0.0
4:00 154.6 F 1.7
6:00 186.8 F 4.7
8:00 206.5 F 8.8
10:00 222.3 F 12.6
12:00 228.4 F 8.9
14:00 237.7 F 11.3
16:00 240.8 F 12.0
18:00 240.2 F 10.9
20:00 253.2 F 17.5
22:00 264.6 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.7
2:00 271.7 F 28.9
3:00 272.4 F 29.8
4:00 272.4 F 29.4

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 210.0 F 0.0
10:00 223.9 F 0.0
15:00 232.1 F 0.0
20:00 227.8 F 0.0
25:00 228.7 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:02:02
CYCLE COMPLETE

08/29/2022 01:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____