
Midmark M11 - v1.0.0
Total Cycles: 4114

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2023 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 197.2 F 3.4
4:00 194.5 F 8.2
6:00 213.6 F 12.6
8:00 224.2 F 8.0
10:00 238.8 F 13.3
12:00 238.6 F 10.7
14:00 236.5 F 8.9
16:00 251.4 F 15.9
18:00 263.5 F 22.9
20:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.6
2:00 271.8 F 28.6
3:00 272.4 F 28.5
4:00 271.6 F 28.5

Min 271.4 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 237.8 F 0.0
10:00 228.5 F 0.0
15:00 227.8 F 0.0
20:00 224.9 F 0.0
25:00 224.9 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

12/13/2023 05:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____