
Midmark M11 - v1.0.0
Total Cycles: 2032

Name: RITTER 1
Sterilizer ID: V2178198
06/16/2022 02:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.8 F 0.0
2:00 137.6 F 1.0
4:00 150.1 F 3.5
6:00 178.2 F 7.3
8:00 197.5 F 12.2
10:00 215.0 F 5.6
12:00 227.9 F 10.2
14:00 233.5 F 8.6
16:00 239.7 F 11.9
18:00 237.8 F 9.9
20:00 251.4 F 16.4
22:00 263.1 F 23.1
24:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.2
2:00 271.7 F 29.0
3:00 272.3 F 29.0
4:00 271.9 F 29.1

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 206.6 F 0.0
10:00 210.8 F 0.0
15:00 216.1 F 0.0
20:00 222.2 F 0.0
25:00 218.8 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

06/16/2022 03:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____